

## Calibration Certificate

<b>Certificate No.:</b>	TZ20180920H	<b>Certificate Date:</b>	September 20,2018
-------------------------	-------------	--------------------------	-------------------

### Reference Instrumentation

Device	Manufacturer	Equipment NO	Accuracy	Calibration Date
Programmable high and low temperature damp heat test chamber	Donguan KeSion precision instrument Co.,LTD.	KS-HW80L	±0.05°C;±3%	June,15,2018
High accuracy RTD thermometer	OMEGA	HH376	±0.15°C	July,5,2018

Tzone Digital Technology Co., Ltd. certifies that the devices mentioned on this certificate have been thoroughly tested, validated and reached performance accuracy specifications over the specific ranges. These device were assembled, tested and calibrated in accordance with its specifications and prior to release for shipment. All reference instruments are calibrated by Shenzhen Academy of Metrology & Quality Inspection, a CNAS (China National Accreditation Service) accredited laboratory, and traceable to NIST according to Multi - Recognition Agreement.

### Product Information

Product Name	Model	Serial No	Production Date	Expiry Date	Quantity
TZONE PDF USB TEMPERATURE DATA LOGGER	TempU02	TZ1809250001-TZ1809250050	9/2018	9/2020	50pcs

### Calibration Results (Environment Conditions: Air Temperature 20.5°C, Relative Humidity 53.0%)

Validation Points	Tolerance	Result
-20°C	±0.5°C	Pass
0°C	±0.5°C	Pass
30°C	±0.5°C	Pass

<b>Test Result Pass/Fail</b>	Pass	<b>Signature of Calibration Technician</b>	<i>Andy Zhang</i>
------------------------------	------	--	-------------------

#### TZONE DIGITAL TECHNOLOGY COMPANY LIMITED

16D, Haiying building, South of Caitian Road, Futian District, Shenzhen City, China. 518033

Tel: 0086-755-82840647 Fax: 0086-755-82840648 Email: sales@tzonedigital.com Web: www.tzonedigital.com